

 Caution — consult accompanying documentation.


To Record an FVC Test

1. Connect spirometer to electrocardiograph.
2. Press .
3. Enter or search for patient data (age, height, gender, race).
4. Select effort and curve types.
5. Press **Start**. Perform effort. Press **Stop**.
6. Accept or reject effort. Repeat from Step 4 until finished.
7. Press **Test Done**.


To Calibrate the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings > Calibration Settings > Calibrate Spirometer**.
4. Enter transducer lot code and calibration code.
5. Update ambient temperature, humidity, and barometric pressure. (See your local weather reports.)
6. Follow prompts.


To Configure the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings**.
4. Change desired settings.


800 535 6663 USA · For worldwide Technical Support numbers, see spirometry manual.
©2005 · Printed in USA · www.welchallyn.com · Material # 703977 Rev B

 Caution — consult accompanying documentation.


To Record an FVC Test

1. Connect spirometer to electrocardiograph.
2. Press .
3. Enter or search for patient data (age, height, gender, race).
4. Select effort and curve types.
5. Press **Start**. Perform effort. Press **Stop**.
6. Accept or reject effort. Repeat from Step 4 until finished.
7. Press **Test Done**.


To Calibrate the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings > Calibration Settings > Calibrate Spirometer**.
4. Enter transducer lot code and calibration code.
5. Update ambient temperature, humidity, and barometric pressure. (See your local weather reports.)
6. Follow prompts.


To Configure the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings**.
4. Change desired settings.


800 535 6663 USA · For worldwide Technical Support numbers, see spirometry manual.
©2005 · Printed in USA · www.welchallyn.com · Material # 703977 Rev B

 Caution — consult accompanying documentation.


To Record an FVC Test

1. Connect spirometer to electrocardiograph.
2. Press .
3. Enter or search for patient data (age, height, gender, race).
4. Select effort and curve types.
5. Press **Start**. Perform effort. Press **Stop**.
6. Accept or reject effort. Repeat from Step 4 until finished.
7. Press **Test Done**.


To Calibrate the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings > Calibration Settings > Calibrate Spirometer**.
4. Enter transducer lot code and calibration code.
5. Update ambient temperature, humidity, and barometric pressure. (See your local weather reports.)
6. Follow prompts.


To Configure the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings**.
4. Change desired settings.


800 535 6663 USA · For worldwide Technical Support numbers, see spirometry manual.
©2005 · Printed in USA · www.welchallyn.com · Material # 703977 Rev B

 Caution — consult accompanying documentation.


To Record an FVC Test

1. Connect spirometer to electrocardiograph.
2. Press .
3. Enter or search for patient data (age, height, gender, race).
4. Select effort and curve types.
5. Press **Start**. Perform effort. Press **Stop**.
6. Accept or reject effort. Repeat from Step 4 until finished.
7. Press **Test Done**.

To Calibrate the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings > Calibration Settings > Calibrate Spirometer**.
4. Enter transducer lot code and calibration code.
5. Update ambient temperature, humidity, and barometric pressure. (See your local weather reports.)
6. Follow prompts.

To Configure the Spirometer

1. Connect spirometer to electrocardiograph.
2. Press Menu key .
3. Choose **Spirometry Settings**.
4. Change desired settings.

800 535 6663 USA · For worldwide Technical Support numbers, see spirometry manual.
©2005 · Printed in USA · www.welchallyn.com · Material # 703977 Rev B